

附件

实验室安全隐患清单


序号	实验室	条款号	问题事实描述	现场照片
1	1103	6.2.3	实验室安全值日记录表未更新	
2	1115	6.2.3	实验室安全值日记录表未更新	
3	1115	9.5.2	危化品防护柜未锁	
4	1310	6.2.3	实验室安全值日记录表未更新	

5	1314	6.2.3	实验室安全值日记录表未更新	
6	1314	8.1.1	插线板未固定,在地面放置	
7	1314	9.3.5	化学试剂瓶标签不规范	
8	1403	6.2.3	无实验室安全值日记录	
9	1405	8.1.1	插线板未固定,在地面放置(2处)	


10	1405	6.2.2	实验室放有折叠床	
11	1404B	9.6.2	气瓶压力表未归零	
12	1408A	9.5.3	易制爆化学品( 硝酸钠 ) 未双人双锁保存	
13	2103	6.2.3	实验室安全值日记录表未更新	
14	2104	9.5.2	危化品防护柜未锁	

15	2105	6.2.3	实验室安全值日记录表未更新	
16	2106	6.2.2	废旧试剂瓶未及 时上交	
17	2109	6.2.3	实验室安全值日记录表未更新	
18	2113	9.7.2	废液桶标签信息 不完整	
19	2113	8.1.1	使用非新国标插 线板	


20	2113	13.2.3	高压灭菌锅使用 记录不规范	
21	2203	6.2.3	实验室安全值日 记录表未更新	
22	2204	6.2.2	实验台上摆放零 乱	
23	2204	7.2.3	洗眼器未定期检 查	
24	2206	6.2.2	实验台上摆放零 乱	

25	2207	8.1.2	水龙头漏水	
26	2207	9.3.1	危化品清单不规范，且无 2 名管理人员的信息(姓名、联系方式)	
27	3105	6.2.3	实验室安全值日记录表未更新	
28	3107	9.3.2	存放化学试剂瓶的实验台架无挡板	
28	3111	6.2.3	无实验室安全值日记录	

30	3113	8.1.1	插线板未固定 ,在 地面放置	
31	3113	8.1.1	实验室地面电线 裸露	
32	3116	6.2.3	实验室安全值日 记录表未更新	
33	3201	9.6.2	气瓶未有效固定 , 且压力表未归零	
34	3201	9.5.3	易制爆化学品( 硝 酸铵 ) 未双人 双锁保存	

35	3203	6.2.3	实验室安全值日记录表未更新	
36	3203	13.4.4	烘箱无操作规程	
37	3205	6.2.3	实验室安全值日记录表未更新	
38	3207	6.2.3	实验室安全值日记录表未更新	
39	3207	9.3.1	危化品清单不规范，且只有 1 名管理人员的信息 (需要 2 名管理人员信息)	

40	3207	9.5.2	危化品防护柜未锁	
41	3207	13.4.4	烘箱无操作规程	
42	3309	6.2.3	实验室安全值日记录表未更新	
43	3315	6.2.3	实验室安全值日记录表未更新	
44	3315	8.1.2	水龙头漏水	
45	3315	7.3.2	通风柜里堆放杂物	

46	3315	8.2.1	学生实验未穿实验服	
47	4103	6.2.3	无实验室安全值日记录	
48	4111	6.2.3	实验室安全值日记录表未更新	
49	4113	6.2.3	实验室安全值日记录表未更新	
50	4115	6.2.3	实验室安全值日记录表未更新	

51	4115	8.1.1	插线板未固定,在地面放置	
52	4116	9.7.2	废液桶标签不规范	
53	4119	6.2.3	无实验室安全值日记录	
54	4203	6.2.3	实验室安全值日记录表未更新	
55	4205	6.2.3	实验室安全值日记录表未更新	
56	4205	8.1.2	水龙头漏水	